


996505
STATINTL



June 24, 1965

C
Dear Bill:

O
In addition to Option I and II offered in our letter of June 4, 1965, we should like to offer Option III which is described as follows:

STATINTL

P
Option III Viewing microscope with 10X to 20X Zoom eyepiece, interchangeable with a recording camera in place of the standard microscope system described in our 1032T Specifications.

Y
It will be necessary to exercise Option III in the early part of construction of the instrument to utilize it. Incorporation of the device into the instrument must be accomplished at the

STATINTL

Very truly yours,

STATINTL

Manager, Engineering

STATINTL CWZ:pc

cc: 

Declass Review by NGA/DOD

991505

STATINTL



June 10, 1965

STATINTL

Subject: Specification dated June 1, 1965, for a Special Microdensitometer
[Redacted] Type 1032T

Dear Bill:


With respect to the subject specifications, we wish to note that the statement below should appear following the second paragraph, Section 3.3, page 3-2.

"A cover plate on the side of the color separation system housing shall allow access to the purifying filters in front of the photomultiplier tubes. It shall then be possible to change the purifying filters in order to shift the wavelength response that remains within the primary separation response widths. Purifying filters that accomplish this shifting are not supplied. Six screws shall be used to secure the cover plate and the changing process will occupy less than thirty minutes."

Additional copies of this letter are enclosed to accompany the specifications already forwarded.

If you have any questions, we will be pleased to hear from you.

Very truly yours,



Asst. Manager, Marketing

STATINTL

AOM:pc

991505

STATINTL

June 4, 1965

Dear Bill:

As requested after our technical discussions concerning your requirements for a digital trichromatic microdensitometric system, we are pleased to provide our quotation.

During our meeting with your personnel, the previous concept of the instrument required was modified to include the following features. Parity Check, special film stage, color wedge, a special rotary stage (10 in. x 10 in.), a longer "Y" measuring motion, a vernier on a standard circle, and a special set of fixed slits. Changes in instruction manuals used with this equipment will also be necessary.

We quote as follows:

STATINTL
STATINTL

One Special Trichromatic Microdensitometer, Type 1032T with Magnetic Tape Output per the enclosed [redacted] [redacted] Specifications dated June 1, 1965.

In addition, we quote the following options:

Option I Air Puck, (Government furnishes a "clean" air supply)

Option II Recording Camera. Includes a special microscope in place of the microscope described in the 1032T Specifications.

It will be necessary to exercise Option II in the early part of construction of this instrument. Incorporation of this device into the entire assembly must be accomplished at the [redacted] Option I STATINTL may be exercised at any time.

Our quotation is FOB Washington, D.C. area and includes installation and checkout of the equipment. It will be firm for a period of 60 days. Delivery will be in 210 to 240 days. Our terms are net 30 days.

Bill

June 4, 1965

It has been our experience at other installations having similar equipment that a 30-day acceptance period has been adequate time for all necessary tests of the equipment to be conducted and evaluated by acceptance personnel. Therefore, the acceptance period for this instrument will not extend beyond this period.

In the event your clean room facilities are not ready for installation of the equipment at the specified time of delivery, we consider that it would be appropriate to receive 90% of the contract amount to relieve us of the hardship of carrying this amount of inventory during a period of delay over which we have no control. In consideration of this, we would be willing to keep the instrument at our facility for a period of up to 90 days without additional charges. We feel it necessary to request the above because we have found by experience that special environmental facilities can take a great deal of time before they are fully ready for installation and checkout of equipment.

If you have any questions, we will be pleased to hear from you and assure a prompt reply.

Very truly yours,

General Manager

STATINTL

BWW:pc